

## NATIONAL INSTITUTE OF UNANI MEDICINE

Kottigepalya, Magadi Main Road, Bangalore – 560 091 (An autonomous body under Ministry of Ayush, Govt. of India) (Accredited to NABH)



Tel: 080-23584260 Fax: 080-23584180 Date: 06/01/2021.

F.No.115-2/2020-21/Accts/NIUM \2\5

To,

Subject: Invitation of Re-Quotation for Supply of Haemoglobin Meter with Accesssories for Conduct Health Survey Camp: – reg.

Sir,

With reference to the subject mentioned above, Re-Quotations are invited from reputed registered dealers for the supply of Haemoglobin Meter with Accesssories for Conduct Health Survey Camp as detailed below.

Sl.	Name of the Items	Specification /Make	Quantity	Amount	Total	Grant
No			Required.		GST	Total
					Amount	
				(₹)	(₹)	(₹)
01	Hb 301 Analyzer	Hemoglobin Analyzer HemoCue Hb	01 Set			
		301 Analyzer:-				
		System: Hemoglobin Analyzer			1 4 1 N	
		Method: Absorbance measurement of				
		whole blood at Hb/HBO <sub>2</sub> Isobestic				
		Point, based on microcuvette				
		technology.				
		Microcuvette: Should be made of				
		polystyrene plastic and contain no				
		active ingredient.				
		Measuring Range: 0-25.6 g/dL.				
		Measurement: Dual Wavelengths				
		of506nm and 880nm for turbidity				
		compensation.		•		
		Analyzer should be USFDA		1 1 1 1 1		
		Certified.			·	
		Measuring: ≤ 3 seconds				
		Correlation of Analyzer with ICSH				
		reference method with $r = 0.995$				
		Sample Volume: ~10 μL				1 * * * * * * * * * * * * * * * * * * *
		Calibration: The system should be				
		factory calibrated ICSH reference				
		need no further calibration.				
.		Quality Control: Built – in self test.	*			
		Interface: RS232 (printer, PC)				
		Working Temperature: 10-40°C.				
		Storage Temp: Cuvettes: 10-40°C,				
		Storage Temp. Cuvettes, 10-40 C,				

Cont. p. No. 02 Cp5.0]

		Anaiyzer: 0-50°C.			
		Power: Battery operated – 4AA			
		Batteries (1.5V).			
		Power Saver Mode: When operating			
		on battery power, the analyser will		•	
		automatically turn off after five			
		minutes of no use.			
		Manufacturer should have ISO			
		9001:2008 certification.			
		Analyzer Complies with IVD			
		Medical Device Directive 98/79/EC		 	
02	HB 301 Cuvetts (Strips)	HB 301 Cuvetts (Strips)	300 No's	 ,	
03	Lancet	Lancet	300 No's		

The quotations may be submitted in a sealed cover quoting the competitive rate with above specification by clearly mentioning taxes applicable if any., time required for supplying the items with the terms & conditions, Copy of GST

Wherever required Test Certificate should be sent along with relevant document. Incomplete bids are liable for rejection. Late / delayed offer will not be considered. NIUM reserves the right to reject any or all tender without assigning any reason.

If the item is not supplied within stipulated period and not as per specification, NIUM is having right to accept or reject whole lot or part of it.

In case the items are not supplied in time, Penalty of 5% for delay upto 15 days from the last date & 10% beyond 15 days upto one month shall be levied. If the firm fails to supply the items as per Supply Order, the item will be procured from other bidder & the difference amount shall be recovered from the bill of the firm or otherwise.

The envelope containing quotation should be super scribed as "Re-Quotation for Supply of Supply of Haemoglobin Meter with Accesssories for Conduct Health Survey Camp"

The last date for the submission of quotation is: 20 / 01 /2020.

Yours faithfully,

(Nadeem A Siddiqui) Administrative Officer

## **Distribution:**

1.Dr. Arish MK Sherwani, HoD, Dept. of TST, NIUM, for information.

2. Website in-charge for information and to upload quotation in NIUM- website.

3.Concern file.

4. Guard file.