



NATIONAL INSTITUTE OF UNANI MEDICINE

Kottigepalya, Magadi Main Road, Bangalore – 560 091

(An autonomous body under Govt. of India, Ministry of Health & Family Welfare)

Affiliated to Rajiv Gandhi University of Health Sciences

E-mail : niumstores@yahoo.in

Tel : 080-23584260

Fax: 080 -23584180

F.No.4-3/2015-16/NIUM/Admn./ 602

16/09/2015

To

Subject: Invitation of quotation for supply of Boyles Anesthesia Machine – reg

Sirs,


With reference to the subject mentioned above, Quotations are invited from reputed registered dealers for the supply of Boyles Anaesthesia Machine required at Hospital Block of NIUM as per the specification enclosed.

The quotations may be submitted in a sealed cover quoting the competitive rate with specification with Brand by clearly mentioning taxes applicable if any., time required for supplying the item with the terms & conditions.

The envelope containing quotation should be super scribed as “**Quotation for the supply of Boyles Anesthesia Machine**”.

The last date for submission of quotation is: **28/09/2015**.

Yours faithfully,


[B. N. Vidya Shankara]
Dy. Director (Admn.)

Specification of Boyles Anesthesia Machine

- ❖ **Boyles Anesthesia Machine** economy model with all accessories, fully SS (complete set)
 - ❖ Anesthesia machine with exclusive nitrous oxide locking system for the highest level of patient's safety
 - ❖ LCP's "ISF" should work as a true fail safe system to allow flow of nitrous oxide only in presence of oxygen at Rota meter level, locks the flow of nitrous oxide in the absence of flow of oxygen through its Rota meter.
 - ❖ Should have tubular, rigid steel section
 - ❖ Provided with sufficient space at the top level of lower drawer to place ventilators, guarded table top with a stainless tray
 - ❖ Instrument tray provided at eye level for keeping monitoring equipments
 - ❖ Should have attachments-
 - ❖ Rota meters : 230 mm long accurately calibrated rotating bobbin flow meters
 - ❖ Oxygen 100ml/min to 8 lpm
 - ❖ Nitrous oxide 200ml/min to 12 lpm
 - ❖ Ether vaporizer
 - ❖ Emergency unit: emergency oxygen flush button is provided at the back bar & guarded to avoid accidental
 - ❖ Passing flow range 35 lpm calibrated at 60 psi (+5%-5%)
 - ❖ Standard Magill's circuit consisting of bag mount- 01 no, corrugated hose 01 no, angle mount 01 no, catheter mount 01 no, face mask 01 no, rebreathing bag 01 no and key spanner 02 no.
-