



NATIONAL INSTITUTE OF UNANI MEDICINE

Kottigepalya, Magadi Main Road, Bangalore – 560 091
(An autonomous body under Ministry of Ayush, Govt. of India)
(Accredited to NABH)



Tel : 080-23584260

Fax : 080 -23584180

Date. 20/11/2020.

F.No.115-2/2020-21/Accts/NIUM/955

To,

Subject: Invitation of Quotation for Supply of Whole Body UV Therapy Unit (Square Shape) for Male & Female Ilaj bit Tadbeer Unit, NIUM: – reg.

Sir,

With reference to the subject mentioned above, Quotations are invited from reputed registered dealers for the supply of Whole Body UV Therapy Unit (Square Shape) for Male & Female Ilaj bit Tadbeer Unit, NIUM as detailed below.

Sl. No	Name of the Items	Specification /Make	Quantity Required.	Amount (₹)	Total GST Amount (₹)	Grant Total (₹)
01	Whole Body UV Therapy Unit (Square Shape)	Whole Body UV Therapy Unit (Square Shape):- It should be made up of powdered coated medical graded mild steel . Shape of the chamber should be square. Inside and outside of the door of the chamber should be provided with handle and lock. Dimensions of chamber: Height 1.95 m, Length 1.0 m and Width 1.0 m. Chamber should have a solid and durable platform. Chamber should be mounted on castor wheels, which allows easy monement. Top of the Chamber (Unit) should be open for proper ventilation. Inner wall of the chamber should have highly polished mirror type aluminium reflectors. Inner tube panel should be protected with fence. Number of Lamps:- UVA : 12 (100 W tubes made by Philips, Holland) NB UVB : 12 (100 W tubes made by Philips, Holland) Electrical supply requirement : 230V+/- 10V, 50Hz, AC supply. Unit should have equally distributed lamps configuration inside the chamber. Unit should have built in Dosimeter system. UV sensors should be provided inside the chamber to monitor variation in	01 Unit			

Cont. p.No. 02
CP.7.07

	<p>irradiation. Unit should be fitted with Micro computerized electronic LCD touch screen controller which allows user friendly setting joules / time for UVA and milli joules / time for NB UVB tubes. The controller should have panel selection to operate individual panel. Unit should have built in memory system to avoid over dosage. In Unit: Fail-safe watchdog circuitry should be provided to prevent excess irradiation. Unit should be fitted with automatic off system with warning alarm at the end of set irradiation time. Earth leakage circuit breakers (ELCB) should be provided in the unit for maximum safety to the patient and the operator. Cooling fans should be provided to prevent overheating.</p>				
--	---	--	--	--	--

The quotations may be submitted in a sealed cover quoting the competitive rate with above specification by clearly mentioning taxes applicable if any., time required for supplying the items with the terms & conditions, Copy of GST

Wherever required Test Certificate should be sent along with relevant document.

Incomplete bids are liable for rejection. Late / delayed offer will not be considered.

NIUM reserves the right to reject any or all tender without assigning any reason.


If the item is not supplied within stipulated period and not as per specification, NIUM is having right to accept or reject whole lot or part of it.

In case the items are not supplied in time, Penalty of 5% for delay upto 15 days from the last date & 10% beyond 15 days upto one month shall be levied. If the firm fails to supply the items as per Supply Order, the item will be procured from other bidder & the difference amount shall be recovered from the bill of the firm or otherwise.

The envelope containing quotation should be super scribed as **“Quotation for Supply of Whole Body UV Therapy Unit (Square Shape) for Male & Female Ilaj bit Tadbeer Unit, NIUM”**

The last date for the submission of quotation is: 18 / 12 / 2020.

Yours faithfully,


(Suresha)
Accounts Officer

Distribution :

1. Prof. Abdul Nasir Ansari, I/c Regimental Therapy Unit, for information.
2. Website in-charge for information and to upload quotation in NIUM- website.
3. Concern file.
4. Guard file.